

UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Addiese: COMMISSIONER FOR PATENTS P O Box 1450 Alexandra, Virginia 22313-1450 www.wepto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/577,128	04/25/2006	Takashi Horaguchi	060340	8185
23850 7590 60723/2008 KRATZ, QUINTOS & HANSON, LLP 1420 K Street, N.W.			EXAMINER	
			BLATT, ERIC D	
Suite 400 WASHINGTON, DC 20005			ART UNIT	PAPER NUMBER
			3734	
			MAIL DATE	DELIVERY MODE
			07/23/2008	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Application No. Applicant(s) 10/577 128 HORAGUCHI, TAKASHI Office Action Summary Examiner Art Unit Eric Blatt 3734 -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --Period for Reply A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS. WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). Status 1) Responsive to communication(s) filed on 15 April 2008. 2a) This action is FINAL. 2b) This action is non-final. 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213. Disposition of Claims 4) Claim(s) 2 is/are pending in the application. 4a) Of the above claim(s) is/are withdrawn from consideration. 5) Claim(s) _____ is/are allowed. 6) Claim(s) 2 is/are rejected. 7) Claim(s) _____ is/are objected to. 8) Claim(s) _____ are subject to restriction and/or election requirement. Application Papers 9) The specification is objected to by the Examiner. 10) The drawing(s) filed on is/are; a) accepted or b) objected to by the Examiner. Applicant may not request that any objection to the drawing(s) be held in abevance. See 37 CFR 1.85(a). Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152. Priority under 35 U.S.C. § 119 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received.

1) Notice of References Cited (PTO-892)

Paper No(s)/Mail Date _

Notice of Draftsperson's Patent Drawing Review (PTO-948)

Information Disclosure Statement(s) (PTO/SE/08)

Attachment(s)

Interview Summary (PTO-413)
Paper No(s)/Mail Date.

6) Other:

Notice of Informal Patent Application

Application/Control Number: 10/577,128

Art Unit: 3734

DETAILED ACTION

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior at are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claim 2 is rejected under 35 U.S.C. 103(a) as being unpatentable over Mangum (US 5,397,326) in view of Clarke (US 3,763,860).

Regarding claim 2, Mangum discloses a suture thread shifting device (Figure 1) wherein a front edge of the device has a sectional configuration suitable for allowing the front edge to reach a region in an arthroscopic or endoscopic viewing field even when an incision portion of a surface layer of a human body is small. The device is capable of having a suture thread engage the front edge, and in doing so, is capable of shifting the suture thread from a position on the operating surgeon's side to a position remote from the operating surgeon. The surfaces of the bisected portions are smoothly curved. The front edge is bisected. In Figures 1 and 2, it appears that the lengths of the two tips may be different from one another, but as Applicant has noted, the figures are not assumed to be drawn to scale and Mangum does not discuss this matter in the text of the specification.

Clarke discloses a needle 46 (Figure 12) designed to carry a suture. Said needle comprises bisected distal end such that a suture may be "held on needle 49 by being

Application/Control Number: 10/577,128

Art Unit: 3734

wedged into a crotch formed between the needle proper and a protuberance 73 thereon." (Column 5, Lines 60-63) Said bisected end comprises bisected portions of differing lengths (Figure 12). Furthermore, the crotch disclosed in Clarke is constructed so as to be able to catch or disengage the suture thread by rotating the device. It would have been obvious to one of ordinary skill in the art at the time of the invention to modify the apparatus of Mangum by providing the crotch of Clarke wherein the bisected portions are of different lengths since said configuration was a known alternative and its substitution would not have produced unexpected results.

Response to Arguments

Applicant's arguments with respect to claim 2 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, THIS ACTION IS MADE FINAL. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the

Application/Control Number: 10/577,128

Art Unit: 3734

shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Eric Blatt whose telephone number is (571)272-9735. The examiner can normally be reached on Monday-Friday, 9:00 AM-6:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Todd Manahan can be reached on 571-272-4713. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Kevin T. Truong/ Primary Examiner, Art Unit 3734

Eric Blatt 571-272-9735